



F 7.1.5 - MASTER CALIBRATION LOG

SERIAL	DESCRIPTION	LOCATION	CALIBRATION METHOD	ACCEPTANCE CRITERIA	DATE CALIBRATED	RESULTS	NEXT CALIBRATION DATE
4012	GAGE BLOCK 0.25"	QA OFFICE	EXTERNAL – IWCS Scale	+/- 0.001"	10/8/2025	PASS	10/8/2026
4052	GAGE BLOCK 0.50"	QA OFFICE	EXTERNAL – IWCS Scale	+/- 0.001"	10/8/2025	PASS	10/8/2026
2915	GAGE BLOCK 1.0"	QA OFFICE	EXTERNAL – IWCS Scale	+/- 0.001"	10/8/2025	PASS	10/8/2026
16157888	MICROMETER	QA OFFICE	GAGE BLOCKS	+/- 0.001"	5/20/2025	PASS	5/20/2026
12177197	MICROMETER	DIE SHOP	GAGE BLOCKS	+/- 0.001"	5/20/2025	PASS	5/20/2026
171407	MICROMETER	KBA40	GAGE BLOCKS	+/- 0.001"	6/23/2025	PASS	6/23/2026
19080585	MICROMETER	QA OFFICE	GAGE BLOCKS	+/- 0.001"	7/14/2025	PASS	7/14/2026
19080587	MICROMETER	SHIPPING	GAGE BLOCKS	+/- 0.001"	7/14/2025	PASS	7/14/2026
3032	MICROMETER	FINISHING	GAGE BLOCKS	+/- 0.001"	5/20/2025	PASS	5/20/2026
22127734	MICROMETER	QA OFFICE	GAGE BLOCKS	+/- 0.001"	11/25/2024	PASS	11/25/2025
22173624	MICROMETER	KBA57	GAGE BLOCKS	+/- 0.001"	11/25/2024	PASS	11/25/2025
28	COUNTER MISTRAL	FINISHING	VISUAL	+/- 0 PCS	6/23/2025	PASS	6/23/2026
38	COUNTER EXPERTFOLD 110	FINISHING	VISUAL	+/- 0 PCS	6/23/2025	PASS	6/23/2026
42	COUNTER EXPERTFOLD 145	FINISHING	VISUAL	+/- 0 PCS	5/20/2025	PASS	5/20/2026
43	EXPERTFOLD MINI	FINISHING	VISUAL	+/- 0 PCS	12/2/2024	PASS	12/2/2025
44	EXPERTFOLD 110-2	FINISHING	VISUAL	+/- 0 PCS	5/6/2025	PASS	5/6/2026
0	SCALE- FLOOR	FINISHING	EXTERNAL - IWCS Scale	+/- 1 LB	10/1/2025	PASS	10/1/2026
Q7814755	SCALE	FINISHING	EXTERNAL - IWCS Scale	+/- 0.001 LB	10/1/2025	PASS	10/1/2026
Q7814768	SCALE	FINISHING	EXTERNAL - IWCS Scale	+/- 0.001 LB	10/1/2025	PASS	10/1/2026
Q7819661	SCALE	FINISHING	EXTERNAL - IWCS Scale	+/- 0.001 LB	10/1/2025	PASS	10/1/2026
Q7819665	SCALE	FINISHING	EXTERNAL - IWCS Scale	+/- 0.001 LB	10/1/2025	PASS	10/1/2026
N/A	ISO 17025 Certificate	IWCS Scale	N/A	N/A			
AWS-600	SCALE	QA OFFICE	REFERENCE ONLY				
171480	MICROMETER	PRESIDENT	REFERENCE ONLY				
12853	MICROMETER	PLANT MANAGER	REFERENCE ONLY				
12854	MICROMETER	QA OFFICE	REFERENCE ONLY				
22127739	MICROMETER	GENERAL MGR	REFERENCE ONLY				
22127740	MICROMETER	OPERATION MGR	REFERENCE ONLY				
001	25' TAPE MEASURE	QA OFFICE	REFERENCE ONLY				
37	25' TAPE MEASURE	DIE SHOP	REFERENCE ONLY				
41	RULER	DIE SHOP	REFERENCE ONLY				